# 510(K) SUMMARY

[as required by section 807.92(c)]

# **FLIGHT 60 Ventilator**

# 510(k) Number K130171

# Date Prepared (revised):

April 7, 2014

## Applicant's Name:

Flight Medical Innovations Ltd.

13 Hamelacha Street

North Industrial Park

Lod 71520, Israel

Tel: (972) 8-9235111 Fax:(972) 8-9236111

**Company Contact:** 

Yanir Shaked, RAQA Manager yanirs@flight-medical.com

### **Contact Persons:**

Shoshana (Shosh) Friedman, RAC; and

**Bosmat Friedman** 

Address:

1914 J N Pease Pl., Charlotte, NC 28262

Telephone: 704-430-8695 or 704-899-0092

Fax:

704-899-0098

E-mail:

shosh@pushmed.com

 $\underline{bosmat@pushmed.com}$ 

#### Trade Name:

FLIGHT 60 Ventilator

### **Common & Classification Name:**

Continuous Ventilator

#### Classification:

Class II; product code 73 CBK and NOU; regulation 21 CFR 868.5895

### **Classification and Review Panel:**

Anesthesiology

## **Predicate Devices:**

- Flight 60 Ventilator, cleared under K120726, manufactured by Fligh Medical Innovations Ltd
- Vela Ventilator, cleared under K032451, manufactured by Bird Product

## **Device Description:**

The FLIGHT 60 Ventilator is an electrically powered, microprocessor controlled ventilator with the following types of ventilatory support: ACMV Volume Pressure or PRVC, SIMV Volume, Pressure or PRVC, PSV/SPONT mode with Pressure Support and Volume Guarantee, Bi-Level (APRV). It can be pressure flow or time triggered; volume or pressure limited; time, pressure or flow cycled. Manual inflation is possible, and an emergency intake valve allows the patient to pull ambient air into the breathing circuit in the event of a complete loss of supply gas pressure.

The FLIGHT 60 may be powered by external power (100 – 240 VACS or 12 – 1 VDC) or by its two internal Li Ion rechargeable batteries, which power the ventilator for up to 12 hours when fully charged.

The electrical system is comprised of three primary boards: the Main boar (motherboard) which holds the majority of the electronics including the mai CPU and the display CPU, the Power board, which holds the power subsystem and internal communication functions, and the Communication board, which holds internal communication and external communication connectors.

The main component of the pneumatic system is an electrically controlled compressor (pump). This compressor provides a compressed gas source so nexternal air compressor is needed. Additionally, the exhalation valve activated by an electrically controlled proportional solenoid that provides built in PEEP.

A comprehensive alarm system is built-in to alert the user to violations of sesafety limits. The alarm system alerts the care giver by activating the audibliaarm, screen display and the LED indicator.

### **Intended Use:**

The FLIGHT 60 Ventilator is intended to provide continuous or intermitter mechanical ventilation support for the care of individuals who require mechanical ventilation. Specifically, the FLIGHT 60 is applicable for adult an pediatric (i.e., infant, child and adolescent) patients, greater than or equal to 5k (11 lbs).

The FLIGHT 60 Ventilator is a restricted medical device intended for use be qualified, trained personnel under the direction of a physician; it is suitable for use in hospital, sub-acute, emergency room, and home care environments, a well as for transport and emergency response applications..

#### **Performance Data**

The FLIGHT 60 Ventilator meets all applicable device specification requirements for performance testing as identified in the FDA reviewed guidance for ventilators. Verification of compliance with recognized standard has been made to support use of the device for its intended use and in intended environment. Additionally, comparison between the performance of the device for its intended use and in intended environment.

the revised Flight 60 Ventilator (subject of this submission) with its predica

نة
픗
æ
Ë
ı.
$\ddot{c}$
=
<u>•</u>
٦
.≥
3
ō
ш

				1
	Trilogy100 Ventilator	Vela Ventilator	FLIGHT 60 Ventilator	FLIGHT 60-0 <sub>2</sub> (100)
				Ventilator
	Respironics Inc.	Bird Products Corp	Flight Medical Innovations Ltd	Flight Medical Innovations Ltd
	K083526	K032451	K120726	K130171
	CBK	CBK	CBK, NOU	CBK, NOU
	The Respironics Trilogy 100	The TBird VELA Ventilator is intended	The FLIGHT 60 Ventilator is	The FLIGHT 60 Ventilator is
7 j	system provides continuous	to provide continuous or intermittent	intended to provide continuous	intended to provide
•		mechanical ventilatory support for the	or intermittent .mechanical	continuous or intermittent
•	support for the care of	care of individuals who require	ventilation support for the care	mechanical ventilation
	individuals who require	mechanical ventilation. The ventilator	of individuals who require	support for the care of
, r	mechanical ventilatio	is a restricted medical device intended	mechanical ventilation.	individuals who require
-	Trilogy 100 is intended for	for use by qualified, trained personnel	Specifically, the FLIGHT 60 is	mechanical ventilation.
•	pediatric through adult	under the direction of a physician.	applicable for adult and pediatric	Specifically, the FLIGHT 60 is
	patients weighing at least 5	Specifically, the ventilator is applicable	(i.e., infant, child and adolescent)	applicable for adult and
	kg (11 lbs.).	for adult and pediatric patients	patients, greater than or equal to	pediatric (i.e., infant, child and
	The device is intended to be	weighing at least 10 kg (22 lbs.), who	5kg (11 lbs).	adolescent) patients, greater
	used in home,	require the following general types of	60 Ventilate	than or equal to 5kg (11 lbs).
	institution/hospital, and	ventilatory support, as prescribed by	restricted medical device	The FLIGHT 60 Ventilator is a
÷	portable applications such as	an attending physician:	intended for use by qualified,	restricted medical device
	wheelchairs and gurneys,	<ul> <li>Positive pressure ventilation</li> </ul>	trained personnel under the	intended for use by qualified,
	and may be used for both	<ul> <li>Assist/Control, SIMV, CPAP modes of</li> </ul>	direction of a physician; it is	trained personnel under the
	invasive and non-invasive	ventilation	suitable for use in hospital, sub-	direction of a physician; it is
:	ventilation. It is not intended	The ventilator is suitable for use in	acute, emergency room, and	suitable for use in hospital,
ι.	to be used as a transport	institutional and transport settings. It	home care environments, as well	sub-acute, emergency room,
٠. ,	ventilator.	is not intended for use as an emergency	as for transport and emergency	and home care environments,
		medical transport ventilator.	response applications.	as well as for transport and
· 17				emergency response
				applications.
ţnı	tures			1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1
it.	•	Yes		Yes
· ]	•	Yes	•	Yes
-	-	Yes	-	Yes
	Yes	-	_	Yes
	Yes	-	•	Yes

## **Standards**

FLIGHT 60 Ventilator has been tested and shown to be compliant with the forstandards:

IEC 60601-1:1998 +A1:1991+A2:1995	Medical electrical equipment - Part 1: General requirements for safety and essential performance
IEC 60601-1-2:2007	Medical electrical equipment General requirements for basic sate essential performance. Collateral standard. Electromagnetic compatibility. Requirements and tests
IEC 60601-1-8:2006	Medical electrical equipment Part 1-8: General requirements safety and essential performance Collateral standard: General requirements, tests and guidance for alarm systems in medical equipment and medical electrical systems
IEC 60601-2-12:2001	Medical electrical equipment Part 2-12: Particular requirement the safety of lung ventilators Critical care ventilators
ASTM F 1246-91	Standard Specification for Electrically Powered Home Care Vent Part 1-Positive-Pressure Ventilators and Ventilator Circuits

### Conclusion

Verification and validation activities were conducted to establish the perfector characteristics of the FLIGHT 60 Ventilator. All testing demonstrated that the 60 Ventilator met required design verification criteria and has acceptable perfewhen used in accordance with its labeling. The device's intended use, o principles, ventilation modes and performance parameters are comparable referenced predicate devices. Therefore, the FLIGHT 60 Ventilator is subsequivalent to the predicate devices.



Food and Drug Administration 10903 New Hampshire Avenue Document Control Center - WO66-G609 Silver Spring, MD 20993-0002

April 8, 2014

Flight Medical Innovations, Ltd C/O Ms. Soshana Friedman, President Push-Med LLC 1914 J.N. Pease Place Charlotte, NC 28262

Re: K130171

Trade/Device Name: FLIGHT 60 Ventilator Regulation Number: 21 CFR 868.5895 Regulation Name: Continuous Ventilator

Class: II

Product Code: CBK Dated: March 18, 2014 Received: March 19, 2014

#### Dear Ms. Friedman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <a href="http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm">http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm</a>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <a href="http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm">http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm</a> for the CDRH's Office

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

of Surveillance and Biometrics/Division of Postmarket Surveillance.

Sincerely yours,



Erin I. Keith, M.S.
Acting Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

## INDICATIONS FOR USE STATEMENT

510(k) Number (if known):K130171					
Device Name:					
FLIGHT60® Ventilator					
Indications for Use:					
The FLIGHT 60 Ventilator is intended to mechanical ventilation support for the omechanical ventilation. Specifically, the pediatric (i.e., infant, child and adolesce (11 lbs).  The FLIGHT 60 Ventilator is a restricted qualified, trained personnel under the ouse in hospital, sub-acute, emergency reas for transport and emergency response.	care of ind FLIGHT 6 ent) patiend d medical direction of boom, and l	lividuals who require to is applicable for adult and tests, greater than or equal to 5kg device intended for use by of a physician; it is suitable for home care environments, as well			
Prescription Use <u>X—</u> (Per 21 CFR 801 Subpart D)	OR	Over the Counter Use (21 CFR 801 Subpart C)			
(PLEASE DO NOT WRITE BELOW THIS LI	INE -CONTIN	UE ON ANOTHER PAGE IF NEEDED)			
Concurrence of CDRH, Offi	ice of Device	Evaluation (ODE)			
	Anya C. Harry -5 2014.04.08 15:03:12 -04'00'				